Service Quotation Number 11084899

PLEASE REFER TO THIS NUMBER WHEN ORDERING

Date 09/19/2017



Pfeiffer Vacuum Inc. 24 Trafalgar Square

Nashua, NH 03063

Tel : 1-800-248-8254 Contact Customer **TRI Univeristy Mezzon Facility** RAHIM ABASALTI 604 222-7652 dba TRIUMF rahim@triumf.ca 4004 Wesbrook Mall Vancouver, BC V6T 2A3 RMA# 13055970 Quote Valid To: 10/19/2017 PO# TR3039905 If you would like to proceed with the Repair, please send revised PO to Suzette Miller smiller@pfeiffer-vacuum.com P 800 248-8254 F 603 578-6550 PRICING IS IN USD Ext. Price Item Description Qty Unit Price 1.0 PUMP ACP28 V2 manual on/off A 2,487.80 2,487.80 1 Catalog # V6SATSFAMF **SERIAL #AC664383** FINDINGS: IN ADDITION TO STANDARD REFURBISHEMENT, UNIT REQUIRES REPLACEMENT OF STATOR (SEIZED) AND BEARING BLOCK (OLD STYLE). ALSO NEED TO ADD JUMPER (MISSING) UNIT REQUIRES: MAINTENANCE KIT **BEARING BLOCK** STATOR JUMPER Repair Estimate shown includes the original \$500.00 1-Year Warranty on parts repaired/replaced Pricing does not reflect shipping charges (shipping charges are extra).

Order Total	\$	2,487.80
Surcharges Taxes (0.00%) Sub Total	\$ \$ \$ \$	2,487.80 0.00 0.00 2,487.80

Quotation Number 11084899

Company Name TRI Univeristy Mezzon Facility



24 Trafalgar Square Nashua, NH 03063 Tel : 1-800-248-8254

Item Description

Qty Unit Price

Ext. Price

If you choose not to proceed with the repair, there will be an evaluation fee assessed. This estimate is contingent upon final test & assembly. Return freight charges are not included.

Terms for delivery and payment:

Delivery FCA Origin Payment Net 30

This quotation includes all of the terms and conditions on both the face and attached documents hereof. The attached detail terms and conditions of sale Pfeiffer document version (Form #14-007) are incorporated by reference as fully as if set forth on the face of this form. This offer may be accepted only on the exact terms set forth herein and no other agreement in any way modifying any of said terms and conditions shall be accepted or binding on Pfeiffer unless accepted in writing. Quoted by: Suzette Miller